

Siemens Healthcare GmbH, HC DI XP MK IN , Allee am Röthelheimpark 2,
91052 Erlangen

<To the person in charge of the unit where the SIEMENS
product is operated, and the administrative head of
organization>


Contact person of ADD MY INFORMATION
the Regional Unit
Department

Telephone
Fax
E-Mail

Date

Customer Safety Advisory Notice

To all users of the Siemens Mammomat Inspiration
with Biopsy functionality



Re: Software Issues on your Mammomat Inspiration system with software version VB30 and Biopsy functionality

Dear Customer,

This letter is to inform you about a potential issue with your the Siemens Mammomat Inspiration running on the software version VB30 and with a Biopsy option.

When do these issues occur and what are the potential risks?

Siemens has become aware of potential errors that may occur while performing a biopsy. The system may freeze and the biopsy functionality may become not available in the following situations:

1. Drop of the compression force, e.g. this may occur due to a patient sudden movement during the active targeting phase.
2. Or use of the inversion tool while the scout image is active during the targeting phase

As a result of a system freeze, the biopsy may have to be repeated.

Siemens Healthcare GmbH
Management: Bernhard Montag, Chairman;
Thomas Rathmann, Michael Reitermann

Allee am Roethelheimpark 2
91052 Erlangen
Germany

Tel.: +49 (9131) 84 0
siemens.com/healthcare

Chairman of the Supervisory Board: Siegfried Russwurm
Registered office: Munich, Germany; Commercial Registry: Munich, HRB 213821
WEEE-Reg.-No. DE 64872105

What steps can the user take to avoid the potential risk of this issue?

- 1.) When the error message is displayed, the user is advised not to close this window but to cancel the active target first and then close the error message window. After these steps are completed the user can start all over.
- 2.) It is advised not to use the inversion tool while the scout image is active.

In case of system freeze the acquisition workplace must be rebooted.

How will the issue finally be resolved?


Siemens will release a field modification to resolve this issue as soon as possible. The new software will be provided to the affected customers free of charge. A Siemens service engineer will contact affected customers to schedule implementation of the modification for your system.

We appreciate your understanding and cooperation with this field safety notice and ask you to immediately instruct your personnel accordingly. Please ensure that this field safety notice is placed in the system's instructions for use. Your personnel should maintain awareness until the modification has been implemented.

If you have sold this device/equipment and it is no longer in your possession, we kindly ask that you forward this safety notice to the new owner of this device/equipment. Please inform us about the new owner of the device/equipment.

We thank you for your cooperation.

Sincerely yours,



Acknowledgement of receipt

Customer address:

I hereby confirm as the owner / responsible operator of the Mammomat Inspiration with the Serial number _____(optional) that I received the following document:

Customer Safety Advisory Notice

Re: Software Issues on your Mammomat Inspiration system with software version VB30 and Biopsy functionality

Place, Date _____

Name _____

Signature _____