



رقم المحفوظات: ٣٨٢٥
رقم الصادر: ١٤/١/٢٧٤٢
بيروت، في: ٢٢ كانون الثاني ٢٠١٢

جانب نقيب المستشفيات الخاصة في لبنان

الموضوع: إشعار بمتابعة جهاز طبي

Ligating Clips, Hem-o-lok Large Polymer Ligating Clips

الجهاز المعني بالمتابعة:

- Ligating Clips, Hem-o-lok Large Polymer Ligating Clips
- Trade Mark: Teleflex Medical
- Local Representative:

بناء على التقرير الصادر عن وكالة ال FDA

الذي يحذر فيه من استعمال الصنف المذكور أعلاه بسبب وجود خلل في التوضيب مما قد يعرض الصنف الوارد أعلاه للتلوث، نرجو منكم تعميم هذه النشرة على جميع المستشفيات المعنية.

مرفق ربطاً:


- التقرير الصادر عن وكالة ال FDA

يبلغ:

دائرة البرامج والمشاريع

المستشفيات الحكومية

المحفوظات



وثيقة مطابقة للأصل
تمت في ٢٦ كانون الثاني ٢٠١٢
رئيس قسم الدروس والنضاب
والملاقات الصحية والتولية
عبد ضومط

U.S. Food & Drug Administration

Enforcement Report - Week of October 24, 2012

1

[FDA Home](#) [Enforcement Reports](#)

[Back to Previous Page](#)

Product Detail

Product Description	Hem-o-lok Large Polymer Ligating Clips, REF #544240, Rx only, Sterile, Teleflex Medical, Research Triangle Park, NC 27709. Teleflex Medicals Hem-o-lok polymer ligating clips are single use, sterile, non-absorbable medical devices that have been designed to ligate vessels and tissue structures during various types of surgical procedures. Vessels and tissue structures in the range of 0.5 to 16 mm can be effectively ligated with Hem-o-lok clips.
Recall Number	Z-0078-2013
Classification	Class II
Code Info	Catalog No. 544240, Lot #01E1200585
Product Distributed Qty	126 units
Reason For Recall	Product is being recalled due to the possibility that the tray may contain pinholes, compromising product sterility.

Event Detail

Event Id	63365
Product Type	Devices
Status	Ongoing
Recalling Firm	Teleflex Medical
City	Research Triangle Park
State	NC
Country	US
Voluntary / Mandated	Voluntary: Firm Initiated
Recall Initiation Date	2012-09-25
Initial Firm Notification of Consignee or Public	Letter
Distribution Pattern	Nationwide Distribution including the state of Nevada.

Links on this page:

- <http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdomain>
- <http://www.addthis.com/bookmark.php>
- <http://www.fda.gov/default.htm>
- <http://www.fda.gov/Safety/Recalls/EnforcementReports/default.htm>

Note: If you need help accessing information in different file formats, see [Instructions for Downloading Viewers and Players](#).

- [Accessibility](#)
- [Contact FDA](#)
- [Careers](#)
- [FDA Basics](#)
- [FOIA](#)
- [No Fear Act](#)
- [Site Map](#)
- [Transparency](#)
- [Website Policies](#)

U.S. Food and Drug Administration
 10903 New Hampshire Avenue
 Silver Spring, MD 20993
 Ph. 1-888-INFO-FDA (1-888-463-6332)
 Email FDA

[USA.gov](http://www.usa.gov)



- [For Government](#)
- [For Press](#)
- [Combination Products](#)
- [Advisory Committees](#)