

22<sup>nd</sup> October 2014

## **URGENT FIELD SAFETY NOTICE – PRODUCT RECALL**

### **GW01 Gastrostomy Guide Wire**

---

Attention: All Nursing, Medical, and Technical staff using this product;

#### **Reason for the Recall Notice**

The safety and performance of these devices have not been properly evaluated and so the CE marking of the device is not valid. We therefore advise that the product is recalled from the market.

#### **Details of Device**

Medicina GW01 Gastrostomy Guide Wire



All batch/lot numbers of the Medicina GW01 Gastrostomy Guide Wire are to be recalled

#### **Action to be taken**

If you hold any stock of Medicina GW01 Gastrostomy Guide Wire could you contact Medicina using the details below and either return to Medicina or Medicina will make arrangements for collection.

**Contact**

Carol Cole  
Quality Assurance  
Unit 2  
Rivington View Business Park  
Station Road  
Blackrod  
Bolton  
BL6 5BN  
Email: [carol@medicina.co.uk](mailto:carol@medicina.co.uk)

Tel: + 44 (0) 1204 695050

Fax: +44 (0) 1204 697755

Medicina confirms that this notice has been notified to the appropriate Regulatory Agency.