

14th November 2014

URGENT FIELD SAFETY NOTICE – PRODUCT RECALL

Medicina Manually Retractable Safety Syringe MRXX & MR05L Manually Retractable Safety Syringe with extended barrel.

Attention: All Nursing, Medical, and Technical staff using this product;

Reason for the Recall Notice

The safety and performance of these devices have not been properly evaluated and so the CE marking of the device is not valid. We therefore advise that the product is recalled from the market.

Details of Device

Medicina Manually Retractable Safety Syringe MRXX (all sizes)



All batches are to be recalled and all sizes of the MRXX including the MR05L Manually Retractable Safety Syringe with extended barrel.

Action to be taken

If any you hold any stock of Medicina MRXX Manually Retractable Safety Syringe we advise that the following actions are taken:

1. Contact Medicina using the details below and either return to Medicina or Medicina will make arrangements for collection.

Contact

Carol Cole
Quality Assurance
Unit 2
Rivington View Business Park
Station Road
Blackrod
Bolton
BL6 5BN
Email: carol@medicina.co.uk

Tel: +44 (0) 1204 695050
Fax: +44 (0) 1204 697755

Medicina confirms that this notice has been notified to the appropriate Regulatory Agency.



FAX BACK FORM

Recall Notice – November 2014

Medicina Manually Retractable Safety Syringe MRXX

Please fax back to: Carol Cole, Quality Assurance, and **Fax:** +44 (0) 1204 697755

We hereby certify that we have received, read, and distributed this Product Recall Notice to anyone it may concern within our organisation.

Hospital/Depot:

Name:

Department:

Function:

Date:

Print name.

Signature: