



[FDA Home](#)³ [Medical Devices](#)⁴ [Databases](#)⁵


Class 2 Device Recall LатарJET EXPERIENCE Combo Screw Driver

[510\(K\)](#)⁶ | [DeNovo](#)⁸ | [Registration & Listing](#)⁹ | [Adverse Events](#)¹⁰ | [Recalls](#)¹¹ | [PMA](#)¹² | [HDE](#)¹³ | [Classification](#)¹⁴ | [Standards](#)¹⁵ | [CFR Title 21](#)¹⁶ | [Radiation-Emitting Products](#)¹⁷ | [X-Ray Assembler](#)¹⁸ | [Medsun Reports](#)¹⁹ | [CLIA](#)²⁰ | [TPLC](#)²¹

[New Search](#)

[Back to Search Results](#)

Class 2 Device Recall LатарJET EXPERIENCE Combo Screw Driver



Date Initiated by Firm	May 12, 2017
Create Date	June 20, 2017
Recall Status¹	Open ³ , Classified
Recall Number	Z-2615-2017
Recall Event ID	<u>77252</u> ²³
510(K)Number	<u>K110763</u> ²⁴ <u>K091694</u> ²⁵ <u>K083096</u> ²⁶
Product Classification	<u>Screw, fixation, bone</u> ²⁷ - Product Code <u>HWC</u> ²⁸
Product	DePuy Mitek LатарJET EXPERIENCE-Combo Screw Driver Product Code: 288211 Product Usage: The Latarjet Cortical Screw set is intended to treat recurrent shoulder instability by supporting the anteroinferior glenoid with a bony graft
Code Information	GTN: 01)10886705026890, Lot codes: 16D02, 16E01, 16J01, 17A01, 17B01, 17B02
Recalling Firm/Manufacturer	DePuy Mitek, Inc., a Johnson & Johnson Co. 325 Paramount Dr Raynham MA 02767-5199
For Additional Information Contact	SAME 508-880-8100
Manufacturer Reason for Recall	Combo Screw Driver (Product Code 288211) tip has the increased potential to break intra-operatively when being used at an angle off-axis to screw
FDA Determined Cause²	Component design/selection
Action	DePuy Mitek issued recall letter dated May 12, 2017 advising of the problem and requesting return of the product. A Response Form is to be completed and returned. Questions: contact your DePuy Synthes Mitek Sales Consultant or Carolyn Somerville, DePuy Mitek, Inc. (via telephone: 508.828.3647 Fax: 508.828.3762 or via email: DPYUS-MitekFieldActions@its.jnj.com).
Quantity in Commerce	173 units
Distribution	US: AZ, CA, CO, GA, MA, MI, OH, WA Foreign: Austria, Australia, Belgium, France, Germany, Spain, Switzerland, Netherlands, Poland, UK
Total Product Life Cycle	<u>TPLC Device Report</u> ²⁹

¹ A record in this database is created when a firm initiates a correction or removal action. The record is updated if the FDA identifies a violation and classifies the action as a recall, and it is updated for a final time when the recall is terminated. Learn more about [medical device recalls](#)³⁰.

² Per FDA policy, recall cause determinations are subject to modification up to the point of termination of the recall.