

## Important information

### PRODUCT RECALL

Product:	Contact lenses	Contact lens care product
Product type:	<input type="checkbox"/> Hard contact lenses <input checked="" type="checkbox"/> Soft contact lenses <input type="checkbox"/> Thermoplastic <input type="checkbox"/> disposable systems	<input type="checkbox"/> Cleaning product for soft contact lenses <input type="checkbox"/> Cleaning product for hard contact lenses

Trading name: Weflex 55 W-50D; Sport Contrast

Lot no.: 100001680091; 100001680231; 100001679962; 100001680179; 100001391625;  
100001391660

Expiry: 6/1/2017; 6/1/2017; 6/1/2017; 6/1/2017; 6/1/2017; 6/1/2017

Type of danger:  
Possible eye irritation or infection

Symptoms & diagnosis  
In rare cases, eye irritation or eye infection may occur on first use of this product.  
There are no external symptoms.

Measures to be taken / information for the end consumer:

These products should not be issued to end consumers. End consumers must be informed that products already issued will have to be returned. If eye irritation or infection occurs, please consult an ophthalmologist immediately.

Please return the above-listed products bearing the stated lot nos. to us immediately.

Please inform your customers about the matter discussed above.

Head of Quality Management

Safety Representative MD