RÉPUBLIQUE LIBANAISE MINISTÈRE DE LA SANTÉ PUBLIQUE

الجمهوريسة اللبنساني وزارة الصحة العامة

المدير العام

Le Directeur Géné

جانب نقيب المستشفيات الخاصة في لبنان

الموضوع: إشعار بمتابعة جهاز دابي Advanced Perfusion System Platform (APS)

الجهاز المعنى بالرتابعة:

- Advanced Perfusion System Platform (APS)
- Trade Mark: Terumo Cardiovascular Systems Corporation
- Local Representative:

بناء على التقرير الصادر عن وكالة ال FDA

الذي يحذر فيه من استعمال الصنف المذكور أعلاه بسبب وجود خلل في عمل الجهاز، نرجو منكم تعميم هذه النشرة على جميع المستشفيات المعنية.

مرفق ربطاً:

- التقرير الصادر عن وكالة ال FDA



U.S. Food & Drug Administration

Medical & Radiation Emitting Device Recalls

• 1

FDA Home³ Medical Devices⁴ Databases⁵



 $510(k)^7 [Registration \& Listing^6] Adverse \ Events^0 [Recalls^{10}] PMA^{11} [Classification^{12}] Standards^{13} \\ CFR Title 21^{14} [Radiation-Emitting \ Products^{16}] X-Ray \ Assembler^{16} [Medsun \ Reports^{17}] CLIA^{16} [TPLC^{19}] \\ The content of the content of$

New Search

Back to Search Results





Date Posted

November 27, 2012

Recall Number

Z-0436-2013

Product

220/240V AC, Advanced Perfusion System Platform (APS) The Terumo® Advanced Perfusion System 1 is indicated for use for up to 6 hours on the extracorporeal circulation of blood for arterial perfusion, regional perfusion, and cardiopulmonary bypass procedures, when used by a qualified medical professional who is experienced in the operation of this or similar equipment.

Code Information

Catalog number: 801764 and all serial numbers

Recalling Firm/

Terumo Cardiovascular Systems Corporation

6200 Jackson Road

Ann Arbor, Michigan 48103-9586

Reason for

Terumo Cardiovascular System (TCVS) has received reports of a situation where users experienced a total loss of functionality for some System 1 units. The reports indicate that the units went blank and shut down with no sign of power and battery backup did not initiate. The result is all pumps stop, with no safety system functionality, and the battery would not be activated. The user would be

Action

Terumo Cardiovascular Systems sent a Urgent Medical Device Recall Correction letter dated November 14, 2012, to all affected customers. The letter identified the product the problem and the action needed to be taken by the customer. Terumo CVS is alerting all users of Terumo System 1 of the reports of malfunction. Customers were instructed to: 1. Review this Medical Device Safety Advisory. 2. Assure that all users are aware of this notice. 3. Confirm receipt of this communication by faxing, or emailing the attached Customer Response Form to the fax number/email address indicated on the form. We encourage you to contact us with any questions or concerns: Terumo CVS Customer Service 1-800-521-2818, Recall Fax 1-734-741-6149 Customer Service Hours: Monday thru Friday. 8 AM - 6 PM ET.

Quantity in Commerce

1647 total units

Distribution

Worldwide Distribution—USA (nationwide) including the states of AL, AR, AZ, CA, CO, CT, DC, DE, FL, GA, HI, IA, ID, IL, IN, KY, LA, MA, MD, ME, MI, MN, MO, MS, MT, NC, ND, NE, NJ, NM, NV, NY, OH, OK, OR, PA, SC, TN, TX, UT, VA, VT, WA, WI, and WV.

Links on this page:

- 1. http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdamain
- 2. http://www.addthis.com/bookmark.php
- 3. http://www.fda.gov/default.htm
- http://www.fda.gov/MedicalDevices/default.htm
- 5. http://www.fda:gov/MedicalDevices/DeviceRegulationandGuidance/Databases/default.htm
- /scripts/cdrh/devicesatfda/index.cfm
- 7. ../cfPMN/pmn.cfm
- 8. ../cfRL/rl.cfm
- 9. ../cfMAUDE/TextSearch.cfm
- 10. ../cfRES/res.cfm
- 11. ../cfPMA/pma.cfm
- 12. ../cfPCD/classification.cfm
- 13. ../cfStandards/search.cfm
- 14. ../cfCFR/CFRSearch.cfm
- 15. ../cfPCD_RH/classification.cfm