

---

## **URGENT FIELD SAFETY NOTICE – PRODUCT RECALL**

### **IVL60 (INS) 60ml Luer Lock IV Syringe for Insulin**

---

Attention: All Nursing, Medical, and Technical staff using this product;

#### **Reason for the Recall Notice**


**22<sup>nd</sup> October 2014**

Medicina Ltd., have been advised that MHRA that the safety and performance of these devices have not been properly evaluated and so the CE marking of the device is not valid.

#### **Details of Device**

Medicina IVL60 (ins) Luer Lock IV Syringe for Insulin

Medicina IVL60 (ins) Luer Lock IV Syringe for Insulin Lot number 12068 expiry date 30/06/2017 are to be recalled



#### **Action to be taken**

If any of Medicina IVL60 (ins) Luer Lock IV Syringe for Insulin are discovered we advise that the following actions are taken:

1. Contact Medicina using the details below and either return to Medicina or Medicina will make arrangements for collection.

**Contact**

Carol Cole  
Quality Assurance  
Unit 2, Rivington View Business Park,  
Station Road, Blackrod, Bolton  
BL6 5BN  
Email: [carol@medicina.co.uk](mailto:carol@medicina.co.uk)  
Tel: 01204 695050  
Fax: 01204 697755

Medicina confirms that this notice has been notified to the appropriate Regulatory Agency.